

		16300496		A
REV	REVISION DESCRIPTION	ECN	DATE	APPROVAL
A	RELEASE TO PRODD.	3-2341	5/2/00	D BELL

HEMISPHERICAL END, C-6 FILAMENT
TUNGSTEN HALOGEN REGENERATIVE CYCLE
MINIATURE TWO-PIN CONSTRUCTION

9.60 ±.01 VOLTS D.C.

1.93 $\pm 5\%$ AMPERES

3300 °K AT FILAMENT CENTER
COILS OUTER DIAMETER USING A
TELESCOPING OPTICAL PYROMETER

42.5 $\pm$ 15% MSCP (LAMP MEASURED
IN A 20" INTEGRATING SPHERE)

35 HOURS AVERAGE (LAB RATED)

$$,780 \pm ,020 \text{ L.C.L.}$$

UNLESS OTHERWISE SPECIFIED DIMENSIONS ARE IN INCHES			DRAWN P. BARAN	DATE 4/00	WELCH ALLYN, INC. SKANEATELES FALLS, NY U.S.A.			
TOLERANCES ON:			CHECKED S. PARKER	4/7/00	TITLE: <i>LAMP SPECIFICATIONS FOR SK-DAB-01331</i>			
BASIC DIMENSION	.XX	.XXX	APPROVED R. DEAN	4/6/00				
UP TO 6	±.02	±.005	REL. FOR PROD. D. BELL	5/2/00				
ABOVE 6 TO 24	±.03	±.010	DO NOT SCALE					
ABOVE 24	±.06	±.015			SCALE		SHEET	OF
ANGLES ± 1/2 DEGREE								