THESE DRAWINGS AND SPECIFICATIONS ARE THE PROPERTY OF WELCH ALLYN, INC. AND SHALL NOT BE REPRODUCED, OR COPIED, OR USED AS A BASIS FOR MANUFACTURE OR SALE OF EQUIPMENT OR DEVICES WITHOUT WRITTEN PERMISSION.

REV	REVISION DESCRIPTION	ECN	DATE	APPROVAL
В	RELEASE IN SAP	1012682	11/12/07	RAD
	ADDED TO STEP 7			

1. LAMP TYPE

REFLECTORIZED LAMP (ELLIPTICAL), C-6 FILAMENT, GAS FILLED

2. RATED VOLTAGE

3.50 ±.01 ∨□LTS D.C.

3. CURRENT

.935 ±5% AMPS

4. COLOR TEMPERATURE

3085 °K AT FILAMENT CENTER CDILS DUTER DIAMETER USING A

TELESCOPING OPTICAL PYROMETER

5. MEAN SPHERICAL CANDLEPOWER

3.23 ±12% MSCP (WIRE TERMINAL LAMP IN A 20" INTEGRATING SPHERE)

6. FOCAL DISTANCE

.50" ±.05" FROM THE REFLECTOR RIM

7. END LIGHT DUTPUT

WITH THE REFLECTOR RIM .50" FROM A .050" DIA. APERTURE AND .50" FROM THE APERTURE TO THE COSINE SURFACE OF THE EG&G COSINE PHOTOMETRIC PROBE, THE LIGHT OUTPUT SHALL BE A MINIMUM OF 375 FOOTCANDLES. THIS MEASUREMENT SHALL BE PERFORMED USING AN EG&G MODEL 550 RADIOMETER/PHOTOMETER SYSTEM EQUIPPED WITH A 550-18 COSINE PHOTOMETRIC PROBE OR EQUIVALENT

8. DESIGN LIFE

300 HOURS AVERAGE (LAB RATED)

9.SPOT SIZE

AT A DISTANCE OF .50° FROM THE REFLECTOR RIM, THE MINOR DIAMETER OF THE OVAL SPOT SHALL BE .080° MINIMUM DIA. WHICH IS LOCATED ON AXIS TO THE REFLECTOR. SPOT SIZE IS DETERMINED BY THE POINT AT WHICH 50%. OF THE CENTER INTENSITY IS REACHED.

UNLESS OTHERWISE SPECIFIED DIMENSIONS ARE IN INCHES			DRAWN PMB		DATE 4/4/94	WEL	_CH ALLYI	V, I	NC.	
TOLERANCES ON:			CHECKED 474794		SKANE	ATELES FALLS,	NY	U.S.A	4.	
BASIC DIMENSION	.××.	.xxx	RAD	4	4/4/94	TITLE:				
UP TO 6	±.02	±.005	APPROVED WKM		4/4/94	LAMP	SPECIFICATIONS FL	JR 710	4-001	
ABOVE 6 TO 24	±.03	±.010	REL. FOR PRO	D.		CODE IDENT	DRAWING NO.		REV.	
ABOVE 24	±.06	±.015	DAB	4	4/5/94	64475	<i>1630028</i>	6-U		B
ANGLES	±.03			r sca	ALE	SCALE		SHEET	DF	