

THESE DRAWINGS AND SPECIFICATIONS ARE THE PROPERTY OF WELCH ALLYN, INC. AND SHALL NOT BE REPRODUCED, OR COPIED, OR USED AS A BASIS FOR MANUFACTURE OR SALE OF EQUIPMENT OR DEVICES WITHOUT WRITTEN PERMISSION.

REV	REVISION DESCRIPTION	ECN	DATE	APPROVAL
B	RELEASE IN SAP ADDED TO STEP 7	1012682	11/12/07	RAD

- | | |
|-------------------------------|--|
| 1. LAMP TYPE | REFLECTORIZED LAMP (ELLIPTICAL), C-6 FILAMENT, GAS FILLED |
| 2. RATED VOLTAGE | 3.50 ±.01 VOLTS D.C. |
| 3. CURRENT | .935 ±5% AMPS |
| 4. COLOR TEMPERATURE | 3085 °K AT FILAMENT CENTER COILS OUTER DIAMETER USING A TELESCOPING OPTICAL PYROMETER |
| 5. MEAN SPHERICAL CANDLEPOWER | 3.23 ±12% MSCP (WIRE TERMINAL LAMP IN A 20" INTEGRATING SPHERE) |
| 6. FOCAL DISTANCE | .75" ±.05" FROM THE REFLECTOR RIM |
| 7. END LIGHT OUTPUT | WITH THE REFLECTOR RIM .75" FROM A .050" DIA. APERTURE AND .50" FROM THE APERTURE TO THE COSINE SURFACE OF THE EG&G COSINE PHOTOMETRIC PROBE, THE LIGHT OUTPUT SHALL BE A MINIMUM OF 543 FOOTCANDLES. THIS MEASUREMENT SHALL BE PERFORMED USING AN EG&G MODEL 550 RADIOMETER/PHOTOMETER SYSTEM EQUIPPED WITH A 550-18 COSINE PHOTOMETRIC PROBE OR EQUIVALENT |
| 8. DESIGN LIFE | 300 HOURS AVERAGE (LAB RATED) |
| 9.SPOT SIZE | AT A DISTANCE OF .75" FROM THE REFLECTOR RIM, THE MINOR DIAMETER OF THE OVAL SPOT SHALL BE .100" MINIMUM DIA. WHICH IS LOCATED ON AXIS TO THE REFLECTOR. SPOT SIZE IS DETERMINED BY THE POINT AT WHICH 50% OF THE CENTER INTENSITY IS REACHED. |

UNLESS OTHERWISE SPECIFIED DIMENSIONS ARE IN INCHES			DRAWN PMB	DATE 3/31/94	WELCH ALLYN, INC. SKANEATELES FALLS, NY U.S.A.			
TOLERANCES ON:			CHECKED RAD	3/31/94				
BASIC DIMENSION	.XX	.XXX	APPROVED WKM	3/31/94	TITLE: LAMP SPECIFICATIONS FOR 7106-001			
UP TO 6	±.02	±.005	REL. FOR PROD. DAB	4/5/94				
ABOVE 6 TO 24	±.03	±.010			CODE IDENT 64475	A	DRAWING NO. 16300287-U	REV. B
ABOVE 24	±.06	±.015			SCALE			SHEET
ANGLES ± 1/2 DEGREE			DO NOT SCALE					