


THESE DRAWINGS AND SPECIFICATIONS ARE THE PROPERTY OF WELCH ALLYN, INC. AND SHALL NOT BE REPRODUCED, OR COPIED, OR USED AS A BASIS FOR MANUFACTURE OR SALE OF EQUIPMENT OR DEVICES WITHOUT WRITTEN PERMISSION.

		16300296		
REV	REVISION DESCRIPTION	ECN	DATE	APPROVAL

1. LAMP TYPE	REFLECTORIZED LAMP (ELLIPTICAL), C-6 FILAMENT, HALOGEN REGENERATIVE CYCLE
2. RATED VOLTAGE	4.25 ±.01 VOLTS D.C.
3. CURRENT	1.060 ±5% AMPS
4. COLOR TEMPERATURE	3000 °K AT FILAMENT CENTER COILS OUTER DIAMETER USING A TELESCOPING OPTICAL PYROMETER
5. MEAN SPHERICAL CANDLEPOWER	3.90 ±12% MSCP (WIRE TERMINAL LAMP IN A 20" INTEGRATING SPHERE)
6. FOCAL DISTANCE	.75" ±.05" FROM THE REFLECTOR RIM
7. END LIGHT OUTPUT	WITH THE REFLECTOR RIM .75" FROM A .050" DIA. APERTURE AND .50" FROM THE APERTURE TO THE COSINE SURFACE OF THE EG&G COSINE PHOTOMETRIC PROBE, THE LIGHT OUTPUT SHALL BE A MINIMUM OF 403 FOOTCANDLES. THIS MEASUREMENT SHALL BE PERFORMED USING AN EG&G MODEL 550 RADIOMETER/PHOTOMETER SYSTEM EQUIPPED WITH A 550-18 COSINE PHOTOMETRIC PROBE
8. DESIGN LIFE	650 HOURS AVERAGE (LAB RATED)
9.SPOT SIZE	AT A DISTANCE OF .75" FROM THE REFLECTOR RIM, THE MINOR DIAMETER OF THE OVAL SPOT SHALL BE .130" MINIMUM DIA. WHICH IS LOCATED ON AXIS TO THE REFLECTOR. SPOT SIZE IS DETERMINED BY THE POINT AT WHICH 50% OF THE CENTER INTENSITY IS REACHED.

UNLESS OTHERWISE SPECIFIED DIMENSIONS ARE IN INCHES			DRAWN PMB	DATE 6/14/94	WELCH ALLYN, INC. SKANEATELES FALLS, NY U.S.A.			
TOLERANCES ON:			CHECKED		TITLE: LAMP SPECIFICATIONS FOR 8106-002			
BASIC DIMENSION	.XX	.XXX	APPROVED		CODE IDENT		DRAWING NO.	
UP TO 6	±.02	±.005	WKM	6/17/94	64475		16300296	
ABOVE 6 TO 24	±.03	±.010	REL. FOR PROD.				REV.	
ABOVE 24	±.06	±.015	DAB	6/23/94				
ANGLES ± 1/2 DEGREE			DO NOT SCALE		SCALE		SHEET OF	