

THESE DRAWINGS AND SPECIFICATIONS ARE THE PROPERTY OF WELCH ALLYN, INC. AND SHALL NOT BE REPRODUCED, OR COPIED, OR USED AS A BASIS FOR MANUFACTURE OR SALE OF EQUIPMENT OR DEVICES WITHOUT WRITTEN PERMISSION.

		16300302		
REV	REVISION DESCRIPTION	ECN	DATE	APPROVAL

1. LAMP TYPE

REFLECTORIZED LAMP (PARABOLIC), C-6 FILAMENT, HALOGEN REGENERATIVE CYCLE
2. RATED VOLTAGE

4.25 ±.01 VOLTS D.C.
3. CURRENT

1.060 ±5% AMPS
4. COLOR TEMPERATURE

3000 °K AT FILAMENT CENTER COILS OUTER DIAMETER USING A TELESCOPING OPTICAL PYROMETER
5. MEAN SPHERICAL CANDLEPOWER

3.9 ±12% MSCP (WIRE TERMINAL LAMP IN A 20" INTEGRATING SPHERE)
6. DIVERGENCE ANGLE

DIVERGENCE 7° ±2° AT GREATER THAN 36" FROM REFLECTOR RIM
7. END LIGHT OUTPUT

A. WITH THE REFLECTOR RIM 36" FROM THE COSINE SURFACE OF THE EG&G COSINE PHOTOMETRIC PROBE, THE LIGHT OUTPUT SHALL BE A MINIMUM OF 90 FOOTCANDLES. THIS MEASUREMENT SHALL BE PERFORMED USING AN EG&G MODEL 550 RADIOMETER/PHOTOMETER SYSTEM EQUIPPED WITH A 550-18 COSINE PHOTOMETRIC PROBE

B. DARK CENTER IN LIGHT SPOT VISUALLY DISAPPEARS AT 2 TO 5" FROM REFLECTOR RIM
8. DESIGN LIFE


650 HOURS AVERAGE (LAB RATED)
9. SPOT SIZE

$2*(d*\text{TAN}(A/2))$ = SPOT DIAMETER IN INCHES

d = DISTANCE FROM REFLECTOR RIM TO TARGET PLANE

A = DIVERGENCE ANGLE

SPOT SIZE IS DETERMINED BY POINT AT WHICH 50% OF CENTER INTENSITY IS REACHED. CALCULATION GOOD FOR GREATER THAN 36" FROM REFLECTOR RIM

UNLESS OTHERWISE SPECIFIED DIMENSIONS ARE IN INCHES			DRAWN	DATE	WELCH ALLYN, INC.			
TOLERANCES ON:			<i>PMB</i>	<i>6/21/94</i>	SKANEATELES FALLS, NY U.S.A.			
BASIC DIMENSION	.XX	.XXX	CHECKED		TITLE:			
UP TO 6	±.02	±.005	APPROVED		<i>LAMP SPECIFICATIONS FOR 8106-003</i>			
ABOVE 6 TO 24	±.03	±.010	<i>WKM</i>	<i>6/21/94</i>	CODE IDENT		DRAWING NO.	REV.
ABOVE 24	±.06	±.015	REL. FOR PROD.	<i>DAB</i>	<i>6/23/94</i>		<i>64475</i>	<i>16300302</i>
ANGLES ± 1/2 DEGREE			DO NOT SCALE		SCALE			SHEET OF