

		16300457		A
REV	REVISION DESCRIPTION	ECN	DATE	APPROVAL
A	DRAWING RELEASE	3-1993	7/15/98	DAB

LENSED END, C-6 FILAMENT, TUNGSTEN-
HALOGEN REGENERATIVE CYCLE

6.00 ±.01 VOLTS D.C.


1.80 $\pm 5\%$ AMPS

2920 °K (USING A TELESCOPING OPTICAL PYROMETER)

9.60 $\pm 15\%$ MSCP (WIRE TERMINAL LAMP
IN A 20" INTEGRATING SPHERE)

.150" NOMINAL DIAMETER MEASURED IN
A TRANSVERSE PLANE TANGENT TO THE LENS

1200 HOURS AVERAGE (LAB RATED)

UNLESS OTHERWISE SPECIFIED DIMENSIONS ARE IN INCHES			DRAWN P. BARAN		DATE 7/13/98		WELCH ALLYN, INC. SKANEATELES FALLS, NY U.S.A.			
TOLERANCES ON:			CHECKED R. DEAN		7/14/98		TITLE: LAMP SPECIFICATIONS FOR 01227			
BASIC DIMENSION	.XX	.XXX	APPROVED D. WITTER		7/15/98					
UP TO 6	±.02	±.005	REL. FOR PROD. D. BELL		7/15/98					
ABOVE 6 TO 24	±.03	±.010	DO NOT SCALE			CODE IDENT 64475		DRAWING NO. 16300457		REV. A
ABOVE 24	±.06	±.015				SCALE				SHEET OF
ANGLES ± 1/2 DEGREE										